



March 6, 2023

Winner Medical Co., Ltd.
Mingni Liu
Regulatory Affairs
Winner Industrial Park, No. 660 Bulong Road,
Longhua District, Shenzhen Guangdong 518109
China

Re: K223232

Trade/Device Name: Procedure mask/Surgical mask/Face mask
Regulation Number: 21 CFR 878.4040
Regulation Name: Surgical Apparel
Regulatory Class: Class II
Product Code: FXX
Dated: February 8, 2023
Received: February 8, 2023

Dear Mingni Liu:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmnmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Katherine D. Kavlock -S

for

Brent Showalter, Ph.D.

Assistant Director

DHT6B: Division of Spinal Devices

OHT6: Office of Orthopedic Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)

K223232

Device Name

Procedure mask/Surgical mask/Face mask

Indications for Use (Describe)

The Procedure mask/Surgical mask/Face mask is intended to be worn to protect both the patient and healthcare personnel from transfer of microorganisms, body fluids, and particulate material. These face masks are intended for use in infection control practices to reduce the potential exposure to blood and body fluids. This is a single-use, disposable device, provided non-sterile.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary

This 510(k) Summary is being submitted in accordance with requirements of Title 21, CFR Section 807.92.

The assigned 510(k) Number: K223232

1. Date of Preparation: October 18, 2022
2. Sponsor Identification

Winner Medical Co., Ltd.

Winner Industrial Park, No.660 Bulong Road, Longhua District, Shenzhen Guangdong, China 518109

Establishment Registration Number: 9616433

Contact Person: Mingni Liu

Position: Regulatory Affairs

Tel: +86-755-28138888-8822

Email: 2346@winnemedical.com

3. Identification of Proposed Device

Trade Name: Procedure mask/Surgical mask/Face mask

Common Name: Surgical Face Mask

Regulatory Information

Classification Name: Mask, Surgical;

Classification: II;

Product Code: FXX;

Regulation Number: 21CFR 878.4040;

Review Panel: General Hospital;

Indications for use:

The Procedure mask/Surgical mask/Face mask is intended to be worn to protect both the patient and healthcare personnel from transfer of microorganisms, body fluids, and particulate material. These face masks are intended for use in infection control practices to reduce the potential exposure to blood and body fluids. This is a single-use, disposable device, provided non-sterile.

Device Description:

The proposed device is a three-layer, single-use, flat-pleated mask. The inner and outer layers of the mask are made of polypropylene nonwoven, and the middle layer is made of polypropylene melt-blown nonwoven. The proposed devices are available in two types, ear loop and tie-on. The ear loops are made of polyester and spandex, and the ties are made of polypropylene nonwoven. The ear loops/ties are used to secure the mask over the users' mouth and face. The nose clip is made of Iron and polypropylene. Users can adjust the nose piece according to the shape of the bridge of the nose, and fix the mask on the bridge of the nose to prevent the mask from falling off. The ear loop masks are available in two size, 17.5×9.5 cm and 14.5×9 cm; the tie-on mask is available in one size, 17.5×9.5 cm. And the colors for the ear loop mask are blue and black; the color for the tie-on mask is blue. Both the ear loop and tie-on masks are available in level 3 masks based on ASTM F2100-19 due to the difference in gram weight of the mask body. The proposed device is provided in non-sterile. All specifications of the proposed device are provided in table 1.

Table 1 Procedure mask/Surgical mask/Face mask Description

Product name	ASTM F2100-19 level	Ear strap type	Size (cm)	Color	Layers	Sterilization
Procedure mask/Surgical mask/Face mask	Level 3	Ear loop	17.5x9.5	Blue	3	Non-sterile
				Black		
			14.5x9	Blue		
				Black		
		Tie-on	17.5x9.5	Blue		

4. Identification of Predicate Device

510(k) Number: K220194

Product Name: Procedure mask/Surgical mask/Face mask

5. Summary of Non-Clinical Test

Non clinical tests were conducted to verify that the proposed device met all design specifications as was same/similar to the predicate device. The test results demonstrated that the proposed device complies with the following standards:

- ASTM F1862/F1862M: 2017 Standard Test Method for Resistance of Medical Face Masks to Penetration by Synthetic Blood (Horizontal Projection of Fixed Volume at a Known Velocity);
- ASTM F2299/F2299M-03 (2017) Standard Test Method for Determining the Initial Efficiency of Material Used in medical Face Masks to Penetration by Particulates using Latex Spheres;
- ASTM F2101: 2019 Standard Test Method for Evaluating the Bacterial Filtration Efficiency (BFE) of Medical Face Mask Materials, Using a Biological Aerosol of Staphylococcus aureus;
- ASTM F2100: 2019 Standard Specification for Performance of Materials Used in Medical Face Masks
- EN 14683:2019+AC: 2019 Annex C Medical face masks - Requirements and test methods
- 16 CFR Part 1610 Standard for the Flammability of Clothing Textiles;
- ISO 10993-5:2009 Biological evaluation of medical device- Part 5: Tests for in vitro cytotoxicity
- ISO 10993-10:2010 Biological evaluation of medical device- Part 10: Tests for irritation and skin sensitization

Table 2 Summary of Performance Testing

Test Methodology	Purpose	Acceptance Criteria	Result
Resistance to Penetration by Synthetic blood	The test was performed in accordance with ASTM F1862/F1862M: 2017 Standard Test Method for Resistance of Medical Face Masks to Penetration by Synthetic Blood (Horizontal Projection of Fixed Volume at a Known Velocity) to evaluate the effectiveness of the test article from possible exposure to blood and other body fluids.	Level 3: No penetration at 160 mmHg	Level 3: Pass at 160mmHg
Particulate Filtration Efficiency	The test was performed in accordance with ASTM F2299/F2299M-03 (2017) Standard Test Method for Determining the Initial Efficiency of Material Used in medical Face Masks to Penetration by Particulates using Latex Spheres to determine the particle filtration efficiency (PFE) of the test article.	Level 3: $\geq 98\%$	Blue mask: Pass at 99.75 % Black mask: Pass at 99.03 %
Bacterial Filtration Efficiency	The test was performed in accordance with ASTM F2101: 2019 Standard Test Method for Evaluating the Bacterial Filtration Efficiency (BFE) of Medical Face Mask Materials, Using a Biological Aerosol of Staphylococcus aureus to determine the bacterial filtration efficiency (BFE) of the test article.	Level 3: $\geq 98\%$	Blue mask: Pass at 99.87 % Black mask: Pass at 98.90 %

Differential Pressure	The test was performed in accordance with EN 14683:2019+AC: 2019 Annex C Medical face masks - Requirements and test methods to determine the differential pressure of the test article.	Level 3:<6.0 mmH2O/cm2	Blue mask: Pass at 4.8 mmH2O/cm2 Black mask: Pass at 3.7 mmH2O/cm2
Flammability	The test was performed in accordance with 16 CFR Part 1610 Standard for the Flammability of Clothing Textiles to evaluate the flammability of the test article.	Class 1	Class 1
Cytotoxicity	The test was performed in accordance with ISO 10993-5 Third edition 2009-06-01 Biological evaluation of medical devices - Part 5: Tests for in vitro cytotoxicity to evaluate the cytotoxicity of the test article.	The viability should be $\geq 70\%$ of the blank. And the 50% extract of the test sample should have at least the same or a higher viability than the 100% extract.	The viability was $\geq 70\%$ of the blank. And the 50% extract of the test sample had a higher viability than the 100% extract. Under the conditions of the study, the proposed device was non-cytotoxic.
Sensitization	The test was performed in accordance with ISO 10993-10 Third Edition 2010-08-01 Biological evaluation of medical	Non-sensitizing	Under the conditions of the study, the proposed device

	devices - Part 10: Tests for irritation and skin sensitization to evaluate the sensitization of the test article.		was non-sensitizing.
Irritation	The test was performed in accordance with ISO 10993-10 Third Edition 2010-08-01 Biological evaluation of medical devices - Part 10: Tests for irritation and skin sensitization to evaluate the irritation of the test article.	Non-irritating	Under the conditions of the study, the proposed device was non-irritating.

6. Clinical Test Conclusion

No clinical study is included in this submission.

7. Technological Characteristics Comparison

Table 3 Comparison of Procedure mask/Surgical mask

ITEM	Proposed Device	Predicate Device K220194	Remark
Product Code	FXX	FXX	Same
Regulation No.	21 CFR 878.4040	21 CFR 878.4040	Same
Class	II	II	Same
Indications for Use	The Procedure mask/Surgical mask is intended to be worn to protect both the patient and healthcare personnel from transfer of microorganisms, body fluids, and particulate material. These face masks are	The Procedure mask/Surgical mask is intended to be worn to protect both the patient and healthcare personnel from transfer of microorganisms, body fluids, and particulate material. These face masks are intended for use in infection control practices to reduce the potential exposure to blood and body fluids. This is a single-use, disposable device, provided non-sterile.	Same

510(k) Summary

	intended for use in infection control practices to reduce the potential exposure to blood and body fluids. This is a single-use, disposable device, provided non-sterile.			
Mask style	Flat pleated	Flat pleated		Same
Design feature	Ear loop / Tie-on	Ear loop / Tie-on		Same
Dimension	<p>Ear loops: Body: 175 mm×95 mm, nose clip: 110mm, Ear-loop: 178mm</p> <p>Body: 145 mm×90 mm, nose clip: 85mm, Ear-loop: 145mm</p> <p>Tie-on: Body: 175 mm×95 mm, nose clip: 110mm, Ear-loop: 900mm</p>	<p>Ear loops: Body: 175 mm×95 mm, nose clip: 110mm, Ear-loop: 178mm</p> <p>Body: 145 mm×90 mm, nose clip: 85mm, Ear-loop: 145mm</p> <p>Tie-on: Body: 175 mm×95 mm, nose clip: 110mm, Ear-loop: 900mm</p>		Same
ASTM F2100 Level	Level 3	Level 1	Level 2	Same
Fluid Resistance	Pass at 160mmHg	Pass at 80mmHg	Pass at 120 mmHg	Same
Particulate filtration efficiency	<p>Blue mask: Pass at 99.87 %</p> <p>Black mask: Pass at 98.90 %</p>	<p>Blue mask: Pass at 96.05%</p> <p>Black mask: Pass at 96.03%</p>	<p>Blue mask: Pass at 98.78%</p> <p>Black mask: Pass at 98.75%</p>	Different
Bacterial filtration	<p>Blue mask: Pass at 99.75%</p>	<p>Blue mask: Pass at 98.25%</p>	<p>Blue mask: Pass at 98.72%</p>	Different

510(k) Summary

efficiency	Black mask: Pass at 99.03 %	Black mask: Pass at 98.25%	Black mask: Pass at 98.73%		
Differential pressure	Blue mask: Pass at 4.8 mmH ₂ O/cm ² Black mask: Pass at 3.7 mmH ₂ O/cm ²	Blue mask: Pass at 3.5 mmH ₂ O/cm ² Black mask: Pass at 3.5 mmH ₂ O/cm ²	Blue mask: Pass at 3.5 mmH ₂ O/cm ² Black mask: Pass at 3.6 mmH ₂ O/cm ²	Different	
Flammability	Class 1	Class 1		Same	
Label/Labeling	Complied with 21 CFR part 801	Complied with 21 CFR part 801		Same	
Materials					
Ear loop	Polyester and spandex	Polyester and spandex			Different
Tie strings	38g /m ² polypropylene nonwoven	38 /m ² polypropylene nonwoven			
Nose clip	Iron and polypropylene	Iron and polypropylene			
Mask body	Outer material	35g/m ² polypropylene nonwoven	Outer material	25g/m ² polypropylene nonwoven	30g/m ² polypropylene nonwoven
	Middle material	33g/m ² polypropylene melt-blown nonwoven	Middle material	25g/m ² polypropylene melt-blown nonwoven	30g/m ² polypropylene melt-blown nonwoven
	Inner material	25g/m ² polypropylene nonwoven	Inner material	25g/m ² polypropylene nonwoven	30g/m ² polypropylene nonwoven
Colors	Blue; Black	Blue; Black		Same	
Biocompatibility					
Cytotoxicity	Under the conditions of the study, the proposed device	Under the conditions of the study, the proposed device was non-cytotoxic.			Same

510(k) Summary

	was non-cytotoxic.		
Sensitization	Under the conditions of the study, the proposed device was non-sensitizing.	Under the conditions of the study, the proposed device was non-sensitizing.	Same
Irritation	Under the conditions of the study, the proposed device was non-irritating.	Under the conditions of the study, the proposed device was non-irritating.	Same
Sterilization			
Method	Non-sterile	Non-sterile	Same

Different - Particulate efficiency

The test result for particulate filtration efficiency for the proposed device is different from the predicate device. However, the test result for the proposed device can meet the requirements of level 3 mask based on ASTM F2100-19. Thus, this difference will not affect the safety and effectiveness of proposed device.

Different - Bacterial filtration efficiency

The test result for bacterial filtration efficiency for the proposed device is different from the predicate device. However, the test result for the proposed device can meet the requirements of level 3 mask based on ASTM F2100-19. Thus, this difference will not affect the safety and effectiveness of proposed device.

Different - Differential pressure

The test result for differential pressure for the proposed device is different from the predicate device. However, the test result for the proposed device can meet the requirements of level 3 mask based on ASTM F2100-19. Thus, this difference will not affect the safety and effectiveness of proposed device.

Different - Materials

The material for the proposed device is the same as the predicate device, both of the proposed device and the predicate device are manufactured by Winner Medical. Level 2 and level 3 masks based on ASTM F2100-19 are made of the same material, except for the difference in the gram weight of the inner, outer and middle layer of the mask. The material of the tie on mask can be covered by the material of the ear loop mask. Therefore, biocompatibility tests were performed on level 2 ear loop blue and black masks, the results does not show any adverse effect. Thus, this difference in the gram weight will

not affect the safety and effectiveness of the proposed device.

8. Conclusion

The conclusions drawn from the nonclinical tests demonstrate that the proposed device is as safe, as effective, and performs as well as the legally marketed predicate device K220194.